



S&S GLOVE CORPORATION

EC DECLARATION OF CONFORMITY

Manufacturers Name and Address: Lot 4, D6 Street, Dat Do I Industrial Park ,
Phuoc Long Tho Commune, Dat Do District,
Ba Ria – Vung Tau Province,, Vietnam

SRN (Single Registration Number): VN-MF•000031848

Authorized Representative Name and Address (if applicable): CMC Medical Devices & Drugs S.L.
C/Horacio Lengo, 18.
CP29006 – Malaga - Spain
Phone: +34.951214054
CMC SRN: ES-AR-000000293
Email: info@cmcmmedicaldevices.com

We, S&S Glove Corporation. hereby declare under our sole responsibility that the product(s) listed below complies with the provisions of the European Medical Device Regulation-MDR 2017/745.

The devices are Class I following Rule I of Annex VIII of the European Medical Device Regulation-MDR 2017/745.

Conformity assessment procedure: following for Class I: Annex II and III of Regulation-MDR 2017/745.

We hereby also declare under our sole that the product below complies with the provisions of (EU) 2016/425 - CAT III.

Product Identification				
Product Name	Model No.	EMDN Code	Basic UDI-DI #	Intended Purpose
Powder Free Nitrile Examination Glove	SSG-N	T01020204	89385409170NCLFT0424229F	Nitrile Examination Glove is intended to be worn by healthcare practitioner to protect patient and

				user from cross contamination
--	--	--	--	-------------------------------

The present declaration is in conformity with European Medical Device Regulation-MDR 2017/745

This declaration is supported by the Quality System Approval to ISO 13485: 2016 and ISO 9001: 2015 issued by SGS United Kingdom Ltd.

Place and Date (dd/mm/yyyy) of Issue :

Ba Ria – Vung Tau, Vietnam (19/12/2023)

Company Representative Signature:



Poppy Farrah Rossa

(QMD Manager)